

● Suction table Timobil



● SERVICE MANUAL

**Dear customer,**

Thank you for choosing a product from the Wassermann range. Wassermann Dental-Maschinen incorporates the highest standards of quality and the latest technology.

In order to enjoy maximum performance and years of trouble-free operation, please read this service manual carefully before you connect this device and start work, and operate the device according to the recommended guidelines. The operation safety and the functionality of this device can only be guaranteed if you follow both the general safety guidelines and the applying laws to prevent accidents as well as the precautions given in this service manual. We are not liable for any damages which occur due to inappropriate usage or faulty operation of this device.



Make sure that anyone using this device has read and understood this service manual.

**Keep this service manual in a safe place where it can be referred to as required at any time.**

**Company address:**

Wassermann Dental-Maschinen GmbH

Rudorffweg 15-17

D-21031 Hamburg, Germany

Phone. : +49 (0)40 / 730 926 – 0 Fax.: +49 (0)40 / 730 37 24

E-mail: [info@wassermann-dental.com](mailto:info@wassermann-dental.com)

**URL: <http://www.wassermann-dental.com>**

## Contents

1 Features.....	4
2 Safety guidelines.....	4
2.1 Safety symbols used in this manual.....	4
2.2 Safety guidelines.....	4
2.3 Duty S6 4 min / 10 min.....	5
2.4 Responsibility for operation or damage.....	5
3 Application.....	6
4 Before starting.....	6
4.1 Transport.....	6
4.2 Installation.....	7
4.3 Storage.....	7
5 Installation / Start-up .....	8
6 Operation.....	8
6.1 General operating instructions.....	8
7 Troubleshooting.....	9
8 Care and maintenance.....	9
8.1 Cleaning.....	9
8.2 Cleaning and/or servicing the chuck.....	10
8.3 Changing the filter.....	10
8.4 Warranty .....	11
8.5 Spare parts.....	11
8.6 Scope of delivery / Accessories.....	11
8.7 Repairs.....	12
9 Technical data.....	12
10 EC Conformity Certificate.....	13
11 Wiring diagram.....	14
12 Spare part diagram.....	15

## 1 Features

- mobile, maintenance- and wear-and-tear-free suction unit for the surgery and the laboratory
- with robust and ergonomic equipment engine
- strong torque, high rotation accuracy, brushless motor
- runs quietly, suitable for continuous operation
- no heat education, no cooling down of the hands
- perspex protection disc in front of the work area
- suction unit can be started via motor handpiece
- high protection via time-delay automatics of the suction unit
- high quality, quickly changed filterbag

## 2 Safety guidelines

### 2.1 Safety symbols used in this manual



#### **Warning!**

This is a warning of risk situations and dangers.

Failure to observe this warning could be life-threatening. These warnings has to be observed.



#### **Information!**

This symbol draws your attention to specific features that has to be observed.

### 2.2 Safety guidelines

Configuring and operating this equipment requires **precise knowledge** and **observance** of the instructions in this **service manual**. The equipment is designed only for its intended application.



**WARNING:**

Servicing and repairs should be carried out only by authorised specialists.



**Warning: Risk of injury!**

**You should change the tool or the chuck only after the motorised handpiece has come to a complete stop. Do not touch the rotating tool.**

### **2.3 Duty S6 4 min / 10 min**

The duty of S6 min 4 min / 10 recommends a loading period of 4 min and a pause of 6 in order to prevent overloading of the control board. Please follow these instructions, otherwise a damage of the control board can be the result and the unit has to be sent to our service department.

### **2.4 Responsibility for operation or damage**

The responsibility for operating the device lies exclusively with the owner or user if said device is incorrectly serviced, maintained or altered by persons not employed by an authorised dealer or if the device is used in a manner contrary to its specified purpose.

The unit has to be maintained and operated in accordance with this user manual. Wassermann Dental-Maschinen GmbH is not responsible for damage arising from the **nonobservance** of these instructions.

Warranty and responsibility provisions contained in the sales and supply conditions of Wassermann Dental-Maschinen GmbH are not extended by these instructions.

### 3 Application

The Timobil is a mobile dust extractor for use in the field and in the laboratory. Its four rollers mean it is easily moved. The electrical cable rolls up automatically into the unit.

The Timobil also has a sturdy handpiece motor.

The speed range is between 1,000 – 40,000 rpm and is infinitely adjustable using a foot pedal.

Turning the casing up to the stop opens and closes the chuck. The unit has overload protection.

The Timobil is a stand-alone dust extractor with an integrated handpiece for use in the laboratory and in the field. It is especially useful for cutting and trimming dental appliances while sitting on a bench stool.

Only use the device for this type of application.

### 4 Before starting

#### 4.1 Transport

Before transporting the unit, ensure that it has been unplugged from the power socket. Make sure that it is packed correctly in order to avoid accidental damage.



Be sure to check for any transport damage when unpacking the goods. Note down any damage if found.

## 4.2 Installation

Open the box, remove the packing materials, and carefully lift out the device and accessories. Check the included accessories.

The accessories are located in the unit, under the filter.

The device has to stand horizontally on a steady and even surface.



Install the device in a place where it will not block the working area and the functionality (take the dimensions into account). Make sure that there are about 10 cm of free space on each side to guarantee air circulation (heat dissipation).



Do not install the unit outdoors or in places without proper ventilation.

Before start-up, be sure the device reaches room temperature.

## 4.3 Storage

If the unit is to be stored for an extended period, protect it from moisture and dust.

## 5 Installation / Start-up


1. Pull the cable out slowly – it will stop at the desired length.
2. Pull the cable out quickly – it will roll back fully onto the drum.
3. **Finally**, insert the power plug into the socket, making sure that the mains and the unit operate on the same voltage.
4. Turn on the green rocker switch.

The unit has an automatic start, that is, the foot control turns the extractor on automatically. The extractor continues running for about 15 seconds after you have finished working.

The toggle switch at the front is used to set **Continuous** operation.

## 6 Operation

### 6.1 General operating instructions

-  All instructions for using the unit, whether in verbal or written form, are based on our own experience and experimentation and can only be regarded as guidelines.



## 7 Troubleshooting

Fault	Cause	Remedy
Little or no extraction	Filter bag full	Change filter bag
No power	Microswitch wrongly adjusted	Reset microswitch. Contact our Service Department,
Motor switches off	Overload protection	Allow unit to cool
Motor does not run	Chuck mechanism open	Close chuck mechanism
Little or no extraction	Filter bag full	Change filter bag



If the above recommendations do not solve the problem, contact your dental depot or our service department.

## 8 Care and maintenance

### 8.1 Cleaning



Disconnect the power plug before starting any maintenance work.

The identification plate has always to be kept in easily legible condition and has not to be removed.



Remove external dirt from time to time with some form of cold cleaner.

Use only cold cleaners to avoid damaging the paintwork or removing the lettering.

The equipment should be cleaned at regular intervals to ensure trouble-free operation. It requires only normal cleaning (sponge, damp cloth, mild detergent) and no further chemical additives.

## 8.2 Cleaning and/or servicing the chuck



### **Warning: Risk of injury!**

**You should change the tool or the chuck only after the motorised handpiece has come to a complete stop. Do not touch the rotating tool.**


1. Open the chuck.
2. Unscrew the plastic tip by turning it manually to the left.
3. Insert the chuck key fully into the chuck.
4. Hold the shaft firmly with an adjustable wrench.
5. Undo the chuck by turning the chuck key to the left.
6. Use the brush to clean the handpiece shaft inside and out. Lubricate the shaft bore outside the chuck with a couple of drops of oil.
7. Insert the chuck by hand.
- 8. Make sure the chuck mechanism is open.**
9. Hold the shaft with the adjustable wrench and use the chuck key to screw the chuck in fully.

## 8.3 Changing the filter

The unit has a one-way filter bag, which is easy to change when it is full. Undo the locks on the right and left and remove the cover. You can now replace the full bag with a new one.

Change the filter regularly to guarantee trouble-free dust extraction. It is time to change the bag if the extraction power becomes noticeably weaker and/or if dust collects in the extractor area.

## 8.4 Warranty

 The warranty period for our equipment is 12 months. If faults occur within the warranty period, contact your dental depot or get in touch directly with our service department.

Your equipment should only be operated in perfect condition. If faults occur which could harm operators or third parties, the unit should not be used until it has been fixed.

This warranty does not cover damage caused by improper use, external mechanical causes, transport damage or interference with the unit by unauthorized persons.

## 8.5 Spare parts

If necessary please contact our service hotline phone:

0049 (0)40 / 730 92 6-0

## 8.6 Scope of delivery / Accessories

<b>Incl. in delivery</b>	<b>Item no.:</b>
Motor handpiece	111121
Rest for handpiece	111103
Protection sheet	149035
Micro fine dust filterbag	- - -

<b>Accessories</b>	<b>Item no.:</b>
Extractor hood	149130
Extractor hood with illumination	149140
Spare filterbag for micro fine dust (dust class M), 1 set = 4 pieces	611010
Spare filterbag for plaster (class 1 / class 2), 1 set = 4 pieces	611018
Storage tray	149010

## 8.7 Repairs



Servicing or repairs to the unit has only to be carried out by qualified technicians. Only original spare parts are to be used. Responsibility for the product is voided if unauthorised persons alter it or if inappropriate components are installed.

## 9 Technical data

<b>Timobil</b>	<b>Artikel Nr.: 149999</b>
Voltage*	220–240 V / 50/60 Hz
Power consumption	1,5 A
Female connector power consumption max	8 A
Output	350 W
W x H x D	530 x 900 x 480 mm
Weight	28,1 kg
Suction capacity	500 m <sup>3</sup> / h (free-blowing)
Control unit capacity	160 W
Handpiece motor capacity	95 W
Speed handpiece motor	1000 – 40 000 min <sup>-1</sup>
Duty	S6 (4 min. / 10 min.)

\*Other voltages upon enquiry.

The noise level of the unit amounts to ≤ 70 dB (A).



The technical data are subject to change without prior notice.

**10 EC Conformity Certificate**

in accordance with 2006/95/EG (low-voltage guidelines) and 2004/108/EG (EMV guidelines) and 2006/42/EG (machinery guidelines)

Manufacturer:	W A S S E R M A N N Dental-Maschinen GmbH Rudorffweg 15 - 17 D-21031 Hamburg	Product description:	Mobile suction table for dental applications
Model:	Timobil Timobil short version	Item no.:	149999 149996
Applicable standards:	DIN EN 61010-1 DIN EN 61000-6-3 DIN EN 61000-6-1 DIN 45635-1 DIN EN 60335-1		

Hiermit wird bestätigt, dass die oben bezeichnete Maschine den genannten EG-Richtlinien entspricht. Diese Erklärung wird ungültig, falls die Maschine ohne unsere Zustimmung verändert wird.

This is to confirm that the above mentioned machine complies with the described EC rules. This declaration becomes invalid if the machine is modified without our approval.

Cette machine est conforme aux normes en vigueur de la Communauté Européene. Cet avis est nul et non avenant si cette machine est modifiée sans notre accord.

Esta máquina, anteriormente mencionada, cumple con los límites requeridos por el reglamento EC. Ahora bien, esta declaración quedará invalidada en caso de realizar modificaciones al aparato sin nuestra aprobación.

Hiermee wordt bevestigd dat bovengenoemde machine voldoet aan de voorgeschreven EU normen. Deze verklaring verliest geldigheid als er zonder onze uitdrukkelijke toestemming wijzigen aan de machine worden aangebracht.

Place, date: Hamburg, 03.05.13

Company stamp:



Signature : \_\_\_\_\_  
Wilfried Wassermann  
(Managing Director)





Wassermann Dental-Maschinen GmbH  
Rudorffweg 15-17 · D-21031 Hamburg  
Tel.: +49 (0)40 730 926-0 · Fax: +49 (0)40 730 37 24  
info@wassermann-dental.com